



## UE DECLARATION OF CONFORMITY

**INMOCLINC, S.L.U.**  
**B88601893**  
**Nº REGISTRO ÚNICO (SNR): ES-MF-000000291**  
**C/ SERRANÍA DE CUENCA, 10**  
**POL. IND. "EL OLIVAR"**  
**28500 ARGANDA DEL REY. MADRID**

Mr. Ángel Moreno González, General Manager of INMOCLINC, S.L.U., Industry of Clinical furniture, **DECLARES** under its responsibility that the device, distributed by INMOCLINC, S.L.U., with the following identification:

Comercial Name:	Examination lamp
Family:	Lamp
Protection Class:	Class II
Referencia:	20102

It is in responsibility for compliance with the requirements of the Directive 2014/35/EC on Low Voltage, Directive 2014/30/EC on electromagnetic compatibility, and complies with the following rules:

- UNE-EN 60598-1:2015+A1:2018      Luminaires. General requirements and tests
- UNE-EN 61000-3-2:2019              Harmonic current emissions ( $\leq 16A$ ) - CEM
- UNE-EN 61000-3-3:2013/A1:2020    Voltage fluctuations and flicker ( $\leq 16A$ ) – CEM

Thus guaranteeing the absence of commitment to the health and safety of persons, provided that the product is used according to its purpose, as well as offering benefits assigned.

INMOCLINC, S.L.U. issues this Declaration of Conformity under the information provided by the manufacturer.

COMMITTS. To establish and maintain a systematical procedure to review experience gained from devices in the post-production phase, using the appropriate means to implement the necessary corrective measures. Likewise,

COMMITTS: To inform the competent authorities on the following facts as soon as it becomes aware of them

- Any malfunction or alteration of the features or benefits, as well as any inadequacy of labeling or instructions for use of a product that might, or might have given rise to the death or serious disruption of the health status of a patient or a user;
- Any reason for technical linked to the characteristics and performance of a device for the reasons stated in the preceding paragraph that has led the manufacturer to systematically withdraw from the market products that belong to the same type.

Date: On June 01, 2022

  
Signature: D. Ángel Moreno González  
General Manager  
INMOCLINC, S.L.U.  
C/ Serranía de Cuenca, 10  
Pol. Ind. El Olivar  
28500 Arganda del Rey  
(Madrid)